



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/632,026	07/31/2003	Jay Lahti	P-11616.00	9661
27581	7590	02/12/2007		
MEDTRONIC, INC. 710 MEDTRONIC PARK MINNEAPOLIS, MN 55432-9924			EXAMINER ALTER, ALYSSA M	
			ART UNIT 3762	PAPER NUMBER

SHORTENED STATUTORY PERIOD OF RESPONSE	MAIL DATE	DELIVERY MODE
3 MONTHS	02/12/2007	PAPER

Please find below and/or attached an Office communication concerning this application or proceeding.

If NO period for reply is specified above, the maximum statutory period will apply and will expire 6 MONTHS from the mailing date of this communication.

Office Action Summary

Application No.

10/632,026

Applicant(s)

LAHTI ET AL.

Examiner

Alyssa M. Alter

Art Unit

3762

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 02 October 2006.
- 2a) ☒ This action is **FINAL**. 2b) ☐ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 1,3-17,19-33 and 35-48 is/are pending in the application.
- 4a) Of the above claim(s) _____ is/are withdrawn from consideration.
- 5) ☐ Claim(s) _____ is/are allowed.
- 6) ☒ Claim(s) 1,3-17,19-33 and 35-48 is/are rejected.
- 7) ☐ Claim(s) _____ is/are objected to.
- 8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☒ The drawing(s) filed on 31 July 2003 is/are: a) ☒ accepted or b) ☐ objected to by the Examiner.
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some * c) ☐ None of:
- ☐ Certified copies of the priority documents have been received.
 - ☐ Certified copies of the priority documents have been received in Application No. _____.
 - ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- | | |
|--|---|
| 1) <input type="checkbox"/> Notice of References Cited (PTO-892) | 4) <input type="checkbox"/> Interview Summary (PTO-413)
Paper No(s)/Mail Date. _____ |
| 2) <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948) | 5) <input type="checkbox"/> Notice of Informal Patent Application |
| 3) <input type="checkbox"/> Information Disclosure Statement(s) (PTO/SB/08)
Paper No(s)/Mail Date _____ | 6) <input type="checkbox"/> Other: _____ |

DETAILED ACTION

Response to Arguments

Applicant's arguments filed July 21, 2006 have been fully considered but they are not persuasive. Therefore the claims stand rejected under *Darby et al.* (US 5,275,620) as stated below.

Double Patenting

The nonstatutory double patenting rejection is based on a judicially created doctrine grounded in public policy (a policy reflected in the statute) so as to prevent the unjustified or improper timewise extension of the "right to exclude" granted by a patent and to prevent possible harassment by multiple assignees. See *In re Goodman*, 11 F.3d 1046, 29 USPQ2d 2010 (Fed. Cir. 1993); *In re Longi*, 759 F.2d 887, 225 USPQ 645 (Fed. Cir. 1985); *In re Van Ornum*, 686 F.2d 937, 214 USPQ 761 (CCPA 1982); *In re Vogel*, 422 F.2d 438, 164 USPQ 619 (CCPA 1970); and, *In re Thorington*, 418 F.2d 528, 163 USPQ 644 (CCPA 1969).

A timely filed terminal disclaimer in compliance with 37 CFR 1.321(c) may be used to overcome an actual or provisional rejection based on a nonstatutory double patenting ground provided the conflicting application or patent is shown to be commonly owned with this application. See 37 CFR 1.130(b).

Effective January 1, 1994, a registered attorney or agent of record may sign a terminal disclaimer. A terminal disclaimer signed by the assignee must fully comply with 37 CFR 3.73(b).

1. Claims 1, 3-17, 19-33 and 35-48 stand provisionally rejected under the judicially created doctrine of obviousness-type double patenting as being unpatentable over claims 1-28 of copending Application No. 10/632058 (US Patent Publication 20050027327 A1). Although the conflicting claims are not identical, they are not patentably distinct from each other because both have a connector assembly for detachably connecting a lead to an implantable medical device, comprising a deflectable connector clip including a first arm, a second arm, the connector clip capable of being deflected from a first position corresponding to a first relative position

Art Unit: 3762

of the first arm and the second arm to a second position corresponding to a second relative position of the first arm and the second arm; and a housing having a first member and a second member, the first member formed to be fixedly engaged with the second member to enclose the connector clip within the housing, wherein the connector clip is positioned within one of the first member and the second member.

This is a provisional obviousness-type double patenting rejection because the conflicting claims have not in fact been patented.

2. Claims 1, 3-17, 19-33 and 35-48 stand provisionally rejected under the judicially created doctrine of obviousness-type double patenting as being unpatentable over claims 1-30 of copending Application No. 10/632028 (US Patent Publication 20050027326 A1). Although the conflicting claims are not identical, they are not patentably distinct from each other because both have a connector assembly for detachably connecting a lead to an implantable medical device, comprising a deflectable connector clip including a first arm, a second arm, the connector clip capable of being deflected from a first position corresponding to a first relative position of the first arm and the second arm to a second position corresponding to a second relative position of the first arm and the second arm; and a housing having a first member and a second member, the first member formed to be fixedly engaged with the second member to enclose the connector clip within the housing, wherein the connector clip is positioned within one of the first member and the second member.

This is a provisional obviousness-type double patenting rejection because the conflicting claims have not in fact been patented.

Claim Rejections - 35 USC § 102

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

(b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.

1. Claims 1, 3-17, 19-33, 35-48 stand rejected under 35 U.S.C. 102(b) as being anticipated by Darby et al. (US 5275620) as previously made of record and stated below.

Darby et al. discloses an implantable lead connector, which is a clip having a top portion and two arms. The clip is disposed within a housing as seen in figure 5. The first position of the arms is before the arms are placed in the housing, the second position is after the arms are placed in the housing and the third position is after the clip engages the lead.

As to claims 1 and 17, Darby et al. does disclose the second end and the first side wall is not adjacent to and engaged against the second side wall when the connector clip is in the second position. As seen in figure 5, there is space between the two respective ends, 24 and 36. Therefore the first side wall and the second side wall are not adjacent and do not engage against each other when the connector clip is in the second position.

As to claims 3-4, 19-20 and 35-36, the cantilevered strips 12 and 14 are preferably formed of a conductive material, such as stainless steel or titanium.

As to claims 7-9, 24-26 and 39-41, as seen in figure 5, the first arm 14 and the second arm 16 have two corresponding flanges, a first flange where the arm 14 rests and a second flange, Both these flanges create recesses in which the corners of the arms are located.

As to claims 13-14, 29-30 and 45-46, the third flange is one of the sealing rings 60 as depicted in figure 5.

Conclusion

Applicant's amendment necessitated the new ground(s) of rejection presented in this Office action. Accordingly, **THIS ACTION IS MADE FINAL**. See MPEP § 706.07(a). Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).


A shortened statutory period for reply to this final action is set to expire **THREE MONTHS** from the mailing date of this action. In the event a first reply is filed within **TWO MONTHS** of the mailing date of this final action and the advisory action is not mailed until after the end of the **THREE-MONTH** shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than **SIX MONTHS** from the date of this final action.


Art Unit: 3762

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Alyssa M. Alter whose telephone number is (571) 272-4939. The examiner can normally be reached on M-F 9am to 4pm.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Angela Sykes can be reached on (571) 272-4955. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.


Alyssa M Alter
Examiner
Art Unit 3762


Primary Examiner
2/1/2 3762